



**FRESENIUS
KABI**

***** **CERTIFICATE OF ANALYSIS** *****

Article: CompoStop Flex2F PLT processing and leuko red.syst
Article number: FD51600
Batch number: 41TF10GA00 Quantity: 10 640
Manufacturing date: 23 June 2022 Expiry date: 2024-05

| Test | Limits | Results |
|-----------------------|-----------------------------------|----------------|
| Configuration | Conform product specification | Conform |
| Physical control | In accordance with specifications | Conform |
| Bacterial Endotoxines | < 20 EU/device | Conform |
| Sterility | Sterile (Ph.Eur.) | Conform |
| Used Filterbatch | 22PB004 | |

| | |
|--|--|
| <u>Batch released by Quality Assurance.</u> | <u>Manufactured by:</u> |
| Date: 17.08.2022 | Fresenius Kabi AG |
| Name: Tykva František | Plant: Fresenius Kabi Horatev CZ s.r.o. |
| Signature:  | Hořátev 104, 289 13 Czech Republic |

Form.no. 17-4269/2