FRESENIUS KABI

***** **CERTIFICATE OF ANALYSIS** *****

Article:	CompoStop Flex2F PLT processing and leuko red.syst		
Article number:	FD51600		
Batch number:	41UH03GA00	Quantity:	9 480
Manufacturing date:	10 August 2023	Expiry date:	2025-07

Test	**Limits**	**Results**
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	22PB015	

Batch released by Quality Assurance.	Manufactured by:
Date: 12.09.2023	Fresenius Kabi AG Plant:
Name: Tykva František	Fresenius Kabi Horatev CZ s.r.o. Hořátev 104, 289 13
Signature:	Czech Republic

Form.no. 17-4269/2